

## CHARACTERIZATION OF SILICON PHOTODIODES FOR DIFFUSE REFLECTANCE SIGNAL EXTRACTION


Pimenta, S.<sup>1</sup>; Carmo, J.P.<sup>1</sup>; Correia, R.G.<sup>1</sup>; Castanheira, E.M.S.<sup>2</sup>; Minas, G.<sup>1</sup>;

<sup>1</sup> Department of Industrial Electronics, University of Minho, Campus de Azurém, Guimarães, Portugal

<sup>2</sup> Centre of Physics (CFUM), University of Minho, Campus de Gualtar, Braga, Portugal  
id4474@alunos.uminho.pt

**KEYWORDS:** *GI cancer, Spectroscopic signals extraction, Photodiodes characterization.*

**ABSTRACT:** Gastrointestinal (GI) cancer is one of most common cause of cancer-related death over the world. Its detection at its earliest stage is crucial to increase the patient survival chances (1). Optical signals extraction and analysis, specifically diffuse reflectance and intrinsic fluorescence, may improve the ability to detect GI dysplasia, once some morphological and biochemical changes on the tissues (related with early cancer progression) can modify these signals' shape and intensity (2, 3). The project under this paper aims to develop a chip sized spectroscopy microsystem for the early detection of GI cancer. This paper presents the characterization of silicon photodiodes (n+/p-epilayer type) fabricated in a standard 0.7  $\mu\text{m}$  CMOS process, with different dimensions. The main goal is to conclude which photodiode must be fabricated for the microsystem implementation, taking into account its capacity to extract the signals in the relevant spectral band (350 nm – 750 nm), not comprising the microsystem dimensions. Figure 1(a) allows concluding that the suitable photodiodes for the detection of those spectroscopic signals are the  $125 \times 125 \mu\text{m}^2$  or  $100 \times 100 \mu\text{m}^2$  active area photodiodes, once their quantum efficiency varies between 20% and 55%, approximately, above 450 nm. In spite of the low photodiodes quantum efficiency (lower than 20%) below 450 nm, the viability of using the  $100 \times 100 \mu\text{m}^2$  photodiode to extract the diffuse reflectance signal, between 350 nm and 750 nm, was proven using a test phantom, representative of a GI tissue – Figure 1(b).



**Figure 1.** (a) Measured spectral response of the fabricated photodiodes; (b) Diffuse reflectance spectra for a test phantom using a calibrated commercial photodiode and the fabricated n+/p-epilayer with  $100 \times 100 \mu\text{m}^2$ . A liquid homogeneous phantom was used, with hemoglobin concentration of 0.5 mg/mL and polystyrene 1  $\mu\text{m}$  beads mass concentration of 0.25%.

## REFERENCES

1. Yu C-C, Lau C, O'Donoghue G, Mirkovic J, McGee S, Galindo L, et al. Quantitative spectroscopic imaging for non-invasive early cancer detection. *Optics express*. 2008;16:16227-39.
2. Brown J, Vishwanath K, Palmer GM, Ramanujam N. Advances in quantitative UV-visible spectroscopy for clinical and pre-clinical application in cancer. *Current opinion in biotechnology*. 2009;20:119-31.
3. Pimenta S, Castanheira E, Minas G. Optical Microsystem for Analysis of Diffuse Reflectance and Fluorescence Signals Applied to Early Gastrointestinal Cancer Detection. *Sensors*. 2015;15:3138-53.